

## § 660.36

### § 660.36 Samples and protocols.

(a) The following shall be submitted to the Office of Biological Product Review Sample Custodian (ATTN: HFB-215), Bldg. 29A, Rm. 1C02, Food and Drug Administration, 8800 Rockville Pike, Bethesda, MD 20892, within 30 days after each routine establishment inspection by FDA.

(1) From a lot of final product, samples from a cell panel intended for identification of unexpected antibodies. The sample shall be packaged as for distribution and shall have at least 14 days remaining in the dating period when shipped to the Center for Biologics Evaluation and Research.

(2) A protocol which shall include the following:

(i) Complete test records of at least two donors of the samples submitted, including original and confirmation phenotyping records.

(ii) Bleeding records or receipt records which indicate collection date, volume, and HBsAg test results.

(iii) Manufacturing records which document all steps involved in the preparation of the product.

(iv) Test results which verify that the final product meets specifications.

(v) Identity test results.

(b) A copy of the antigenic constitution matrix specifying the antigens present or absent shall be submitted to the Director, Center for Biologics Evaluation and Research, at the time of initial distribution of each lot of Reagent Red Blood Cells for detection or identification of unexpected antibodies. Products designed exclusively to identify Anti-A, Anti-A<sub>1</sub>, and Anti-B, as well as products composed entirely of umbilical cord cells, are excluded from this requirement.

(c) Except for umbilical cord samples, whenever a new donor is used, a sample of red blood cells from each new donor used in a cell panel intended for the identification of unexpected antibodies shall be submitted by the manufacturer to the Director, Center for Biologics Evaluation and Research. The sample should contain a minimum volume of 0.5 milliliter of red blood cells.

[52 FR 37450, Oct. 7, 1987, as amended at 55 FR 11013 and 11015, Mar. 26, 1990; 67 FR 9587, Mar. 4, 2002]

## 21 CFR Ch. I (4-1-03 Edition)

### Subpart E—Hepatitis B Surface Antigen

SOURCE: 44 FR 36382, June 22, 1979, unless otherwise noted.

#### § 660.40 Hepatitis B Surface Antigen.

(a) *Proper name and definition.* The proper name of this product shall be Hepatitis B Surface Antigen (HBsAg), which shall consist of a serum or tissue preparation containing one or more subtypes of the Hepatitis B Surface Antigen.

(b) *Source.* The source of the product shall be blood, plasma, serum, or tissue, obtained aseptically from nonhuman primates that have met the applicable requirements of § 600.11 of this chapter, or from human donors whose blood is positive for the Hepatitis B Surface Antigen.

#### § 660.41 Processing.

(a) *Method.* The processing method shall be one that has been shown to yield consistently a specific and potent final product, free of properties which would adversely affect the test results when the product is tested by the methods recommended by the manufacturer in the package insert. The product and all ancillary reagents and materials supplied in the package with the product shall be manufactured in a manner that will reduce the risk of transmitting type B viral hepatitis.

(b) *Ancillary reagents and materials.* All ancillary reagents and materials supplied in the package with the product shall meet generally accepted standards of purity and quality and shall be effectively segregated and otherwise manufactured in a manner that will reduce the risk of contaminating the product and other biological products. Ancillary reagents and materials accompanying the product, which are used in the performance of the test as described by the manufacturer's recommended test procedures, shall have been shown not to affect adversely the product within the prescribed dating period.

(c) *Final container.* A final container shall be sufficiently transparent to permit visual inspection of the contents for presence of particulate matter and increased turbidity. The effectiveness